

Notice of Allowability

Application No.

10/781,487

Examiner

Christopher J. Novosad

Applicant(s)

ARCHULETA ET AL.

Art Unit

3671

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to a telephone call from Mr. Curtis P. Robando (Attorney for Applicants).
2. ☒ The allowed claim(s) is/are 1,2 and 5-8.
3. ☒ The drawings filed on 18 February 2004 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/08), Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☒ Interview Summary (PTO-413), Paper No./Mail Date 020205.
7. ☒ Examiner's Amendment/Comment
8. ☐ Examiner's Statement of Reasons for Allowance
9. ☒ Other Comments on Allowance.

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Mr. Curtis P. Robando, Attorney for Applicants, on February 2, 2005.

The application has been amended as follows:

In the claims:

In claim 1, line 14, the recitation "blade;" has been deleted and --blade having a smooth shearing edge;-- has been inserted in lieu thereof.

COMMENTS ON ALLOWANCE

It is to be noted that the saw blade 50 of the secondary reference to Clark (which was used to read on the "coultter blade" recited in claim 1) has a series of saw teeth 54 and is not a smooth shearing edge. There is no suggestion or motivation to change the saw teeth edge 54 of Clark to a smooth edge since this would destroy the function of the saw blade 50 (it would no longer function as a saw blade). In any event, any modification of Clark would involve an improper modification of a secondary reference (Clark). The edges (unnumbered) of the saw blade 50 above the saw teeth 54 of Clark do not function as shearing edges. It is to be further noted that the other art of record, for example Pettersson '551, Larson '028 and Pratt '815 show various agricultural implements with extensions but fail to show coultter blades as extensions and

Art Unit: 3671

also fail to show each of the opposing side walls comprising an extension that is exterior to the pan and is tapered toward the open end of the socket. With respect to Rich *et al.*, extension member 50 is not a coultter blade but is an elongated bar and, in any event, the opposing side walls of member 28 do not have the extension 50, but rather the extension 50 is attached to a flange 29.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher J. Novosad whose telephone number is 703-308-2246. The examiner can normally be reached on Monday-Thursday 5:30am-4:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Thomas Will can be reached on 703-308-3870. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3671

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Christopher J. Novosad
Primary Examiner
Art Unit 3671

February 2, 2005